

Reclassification of some medical devices

Dr Amanda Craig

Director

Devices Post Market Reforms and Reviews Section

Medical Devices Surveillance Branch

Department of Health and Aged Care, TGA



Australian Government

Department of Health and Aged Care
Therapeutic Goods Administration



Urogynaecological mesh

Surgical mesh

Software, anatomical models, diagnostic imaging

Six kinds of medical devices

Several new and amended classification rules

Software, anatomical models, and diagnostic imaging

01

Provides a diagnosis or screen for a disease or condition

02

Monitor the state, progression, or parameters of a disease or condition

03

Specify or recommend a treatment or intervention

04

Provide therapy through the provision of information

05

Record patient images using energy outside the visible spectrum, used in diagnosis or monitoring of disease or condition

06

Virtual or physical anatomical models, or software used to generate the virtual models, used in diagnosis or monitoring of disease or condition



When is software a medical device?

Intended purpose is critical...

Software is a medical device when the manufacturer *intends* for its product to be used for:

- diagnosis
- prevention
- monitoring
- prognosis & prediction
- treatment
- alleviation

...of disease, injury or disability

It's not based on...

- what colour it is
- what shape it is
- whether or not it looks good
- what it is called
- whether it is novel or not
- what technology is used
- how it is supplied (online, in hardware, cloud portal ...)



Clinical decision support software

Exemption under particular requirements:


- a) intended to be for the sole purpose of providing or supporting a recommendation to a health professional about preventing, diagnosing, curing or alleviating a disease, ailment, defect or injury in persons; and
- b) not intended by its manufacturer to directly process or analyse a medical image or signal from another medical device; and
- c) not intended by its manufacturer to replace the clinical judgement of a health professional in relation to making a clinical diagnosis or decision about the treatment of patients

Sponsors/suppliers **must** notify the TGA of their exempt CDSS devices using the [Notification form: Clinical decision support software exemption](#)



Excluded software

Exclusions are conditional:



Examples of excluded software:

A 'sun smart' app that gives user alerts for UV protection to minimise skin cancer risk.

Examples of software that is not excluded:

A smartphone app intended to provide a direction to adjust the dosage of a prescribed medication.

25 February 2021 – New classification rules commence

01 November 2024 – Transition period ends

ARTG-included devices
Current classification rules

Notification period
Ends 25 August 2021

Transition period

ARTG-included devices
New classification rules

Devices under application
Current classification rules

Notification period
Ends 2 months after the date of ARTG inclusion

New devices for inclusion in the ARTG
New classification rules

Reclassification of certain medical devices

01

Spinal Implantable medical devices

02

Active Implantable medical devices (AIMD)

03

MD administer medicines or biologicals by inhalation

04

MD substances for introduction into body via body orifice or applied to the skin

05

Active medical devices for therapy with diagnostic function

06

MD in direct contact with the heart, CCS, or CNS



1 November 2021 – New
classification rules commence

01 November 2024 –
Transition period ends

ARTG-included devices
Current classification rules

Notification period
Ends 25 May 2022

Transition period

**ARTG-included
devices**
New classification rules

Devices under application
Current classification rules

Notification period
Ends 2 months after the
date of ARTG inclusion

New devices for inclusion in the ARTG
New classification rules

EU MDR Transition Extension

The European Parliament agreed the following key elements:

- extend the transition for qualifying medical devices from 26 May 2024 to:
 - 26 May 2026 for class III implantable custom-made devices
 - 31 December 2027 for class III and implantable class IIb devices
 - 31 December 2028 for non-implantable class IIb and lower risk devices
- remove the “sell-off” deadline of 25 May 2025 from the MDR and IVDR.



Spinal Implantable Medical Devices

<https://www.tga.gov.au/resource/reclassification-spinal-implantable-medical-devices>

Old classification
Class IIb

New Classification
Class III

Schedule 2 Part 3 Classification rule 3.4 (4B)

3.4 Surgically invasive medical devices intended for long term use and implantable medical devices

- (4B) If the device is intended by the manufacturer to be a motion preserving device for the spine (such as a spinal disc replacement), the device is classified as Class III.

Review snapshot

36 GMDNs in the review – 9 additional GMDNs identified through notifications

Active Implantable Medical Devices (AIMD)

<https://www.tga.gov.au/resource/reclassification-active-implantable-medical-devices-aimd>

Old classification
AIMD

New Classification
Class III

Schedule 2 Part 5 Classification rule 5.7

5.7 Active implantable medical devices

- (1) An active implantable medical device is classified as Class III.
- (2) An implantable accessory to an active implantable medical device is classified as Class III.
- (3) An active medical device that is intended by the manufacturer to be used to control or monitor, or directly influence, the performance of an active implantable medical device is classified as Class III.

Review snapshot

All candidates notified and are transitioning

Medical Devices that administer medicines or biologicals by inhalation

<https://www.tga.gov.au/resource/reclassification-medical-devices-administer-medicines-or-biologicals-inhalation>

Old classification
Class I or IIa

New Classification
Class IIa or Class IIb

Schedule 2 Part 5 Classification rule 5.10

5.10 Medical devices that administer medicines or biologicals by inhalation

If a medical device is intended to be used to administer medicines or biologicals by inhalation:

- (a) if the mode of action of the device has an essential impact on the efficacy and safety of the medicines or biologicals—the device is classified as Class IIb; or
- (b) if the device is intended to treat a life threatening condition—the device is classified as Class IIb; or
- (c) if paragraphs (a) and (b) do not apply—the device is classified as Class IIa

Review snapshot

109 GMDNs in the review – 51 additional GMDNs identified through notifications

MD that are substances for introduction into body via body orifice or applied to the skin

<https://www.tga.gov.au/resource/reclassification-medical-devices-are-substances-introduced-human-body-body-orifice-or-applied-skin>

Old classification
Class I or IIa

New Classification
Class IIa, Class IIb

Schedule 2 Part 5 Classification rule 5.11

5.11 Medical devices that are substances to be introduced into the body or applied to and absorbed by the skin

If a medical device is composed of substances, or combinations of substances, that are intended to be:

- (a) introduced into the human body through a body orifice; or
- (b) applied to and absorbed by the skin;

the device is classified as follows:

- (c) if the device is introduced into the nasal or oral cavity as far as the pharynx, or is applied to and absorbed by the skin, and achieves its intended purpose in that cavity or on the skin—Class IIa;
- (d) in any other case—Class IIb.

Review snapshot

95 GMDNs in the review – 54 additional GMDNs identified through notifications

Active medical devices for therapy with diagnostic function

<https://www.tga.gov.au/resource/reclassification-active-medical-devices-therapy-diagnostic-function>

Old classification
Class IIa or Class IIb

New Classification
Class III

Schedule 2 Part 4 subclause 4.2(4)

Subclause 4.2(4):

(4) An active medical device for therapy that includes a diagnostic function the purpose of which is to significantly determine patient management by the device is classified as Class III.

Review snapshot

29 GMDNs in the review – 71 additional GMDNs identified through notifications

Medical Devices in direct contact with the heart, CCS or CNS

<https://www.tga.gov.au/resource/reclassification-active-implantable-medical-devices-aimd>

Old classification
Class IIa

New Classification
Class III

Schedule 2 Part 3 subclauses 3.2 and 3.3

Subclause 3.2(3A):

(3A) If the device is not a reusable surgical instrument and the device is intended by the manufacturer specifically to be used in direct contact with the heart, the central circulatory system or the central nervous system of a patient, the device is classified as Class III.

Subclause 3.3(4)(b):

(4) If the device is intended by the manufacturer:

... (b) specifically, to be used in direct contact with the heart, the central circulatory system or the central nervous system of a patient;

... the device is classified as Class III

Review snapshot

79 GMDNs in the review – 61 additional GMDNs identified through notifications

NEXT STEPS.....

Non-transitioning devices

Assess non-transitioning devices - s41JAs

Affected:

- * Monitor / assess potential supply issues
- * Provide feedback on classification rule - sponsors requested to cancel their entry and cease supply until they have an approved inclusion at the higher classification.
- * Alternatively, TGA can cancel the affected non-transitioning devices

Not affected:

Provide feedback and confirm no need to reclassify
Close out of the review



NEXT STEPS.....

Transitioning devices

- Reminders to be sent for reclassification applications
- Previous ARTG entries cancelled after a decision is made on applications for the higher classification
- Cancellation of devices that fail to submit application for correct class by 1 Nov 24
- Communication with stakeholders including state and territory health departments, Prothesis List, etc
- Monitor / assess potential supply issues



Website references

TGA website	www.tga.gov.au
Notification form: Clinical decision support software exemption	https://consultations.health.gov.au/medical-devices-and-product-quality-division/clinical-decision-support-exemption/
Reclassification of spinal implantable medical devices	https://www.tga.gov.au/resource/reclassification-spinal-implantable-medical-devices
Active Implantable Medical Devices (AIMD)	https://www.tga.gov.au/resource/reclassification-active-implantable-medical-devices-aimd
Medical Devices that administer medicines or biologicals by inhalation	https://www.tga.gov.au/resource/reclassification-medical-devices-administer-medicines-or-biologicals-inhalation
MD that are substances for introduction into body via body orifice or applied to the skin	https://www.tga.gov.au/resource/reclassification-medical-devices-are-substances-introduced-human-body-body-orifice-or-applied-skin
Active medical devices for therapy with diagnostic function	https://www.tga.gov.au/resource/reclassification-active-medical-devices-therapy-diagnostic-function
Medical Devices in direct contact with the heart, CCS or CNS	https://www.tga.gov.au/resource/reclassification-active-implantable-medical-devices-aimd

Therapeutic Goods Administration (TGA)

Exhibition booth No.1

Want to chat with me further? Come visit us.



Therapeutic Goods Administration



Questions?

www.tga.gov.au



Australian Government

Department of Health and Aged Care
Therapeutic Goods Administration