Laboratories Branch

Therapeutic Goods Administration

Owner:	Number: Bio-BEE-Form-42
Author:	Version: 1
Active: 2/07/2021	Review: <qpulse_docreviewdate></qpulse_docreviewdate>
Title: Destarial Endatesia. Apparatis 4K. Kinetia I Al. Destina Assau Caranla Desulta	

Title: Bacterial Endotoxin - Appendix 1K - Kinetic LAL Routine Assay Sample Results

Appendix 1K - Kinetic LAL Routine Assay Sample Results Sheet

Assay ID: <u>19/7/21</u> Operator:

LAL Reagent Water (LRW) Lot Number: 0000830072 LRW Expiry: 1 March 2022

Other Reagent: Pyrosperse Batch# 0000843500 Expiry: 2 October 2021 Use By:

23 July 2021

1 - Product Details

Product Name	Batch Number	Expiry	LIMS Number
Pfizer-Comirnaty	<u>FE3430</u>	31 October 2021	2107002694-R1
Product Concentration	Endotoxin Limit	MVD	Test Dilution
<u>-</u>		<u>2500</u>	<u>1/1000</u>

2 - Product Dilutions

Dilution	Volume of Dilution	Volume of LRW	Volume of Other
For 1/1000 do: 1/50	20uL of vaccine	<u>975uL</u>	<u>5uL</u>
1/20	50uL of 1/50	<u>945uL</u>	<u>5uL</u>
=	=	-1	-
=	=	=	-
Ξ	-	-	-1

Recording of Results

The results are recorded by the WinKQCL software. Transcribe results from the WinKQCL report to assay sheet.

Test CV (%)	Result (EU/ml)
<u>1.23</u>	<u>< 5.00</u>

PPC CV (%)	% PPC Recovery
<u>1.91</u>	<u>92</u>

OFFICIAL

For the sample test to be valid, the PPC recovery must be 50-200% of the expected value and the coefficient of variation (CV) of the sample and PPC duplicates should be < 10%. Many samples do not reach an endpoint and are not assigned a %CV. In these instances, contact the person in charge of the assay.

Validity Criteria	
Were the standard curve validity criteria met?	Yes
Were the sample validity criteria met?	Yes
Was the result within the endotoxin limit?	Yes

This Appendix is used for recording the sample results. Use the SOP (Appendix 8K) for the detailed method.

Notes:

Click or tap here to enter text.